

**PHYSIO  
CONTROL**

## LIFEPAK CR<sup>®</sup> Plus DEFIBRILLATOR

### Make Lifesaving Simple

With over 50 years of innovation, a steadfast commitment to quality and a position as the global leader in defibrillation, Physio-Control brings you the LIFEPAK CR Plus automated external defibrillator. The *CR Plus* is designed specifically for the first person to respond to a victim of sudden cardiac arrest and incorporates the same trusted technology used by more EMS and hospital units around the world than any other brand.

Although not everyone can be saved from sudden cardiac arrest, studies show that early defibrillation can dramatically improve survival rates.



#### Simple to Use

- Simple to turn on
- Simple to find, remove and place electrodes correctly
- Simple to deliver therapy—no shock button to push
- Simple to increase the chance for survival by automatically escalating energy up to 360 joules if needed\*

#### Simple to Own

- The *CR Plus* comes ready to use: Initial purchase includes carry case, extra electrodes, CHARGE-PAK™ battery charger and Ambu® Res-Cue Mask® Kit
- Lowest total cost of ongoing ownership in the AED industry
- Simple transition to EMS teams who also use LIFEPAK products

#### Simple to Maintain

- One of the longest warranties in the industry at 8 years
- Synchronized CHARGE-PAK battery charger and electrode replacement cycle

#### Simply put...

The LIFEPAK CR Plus automated external defibrillator from Physio-Control is the effective, safe and affordable choice.

AED users should be trained in CPR and the use of an AED. LIFEPAK AEDs require a prescription in the U.S. Please consult your physician.



## SPECIFICATIONS

### DEFIBRILLATOR

**Waveform:** Biphasic truncated exponential, with voltage and current duration compensation for patient impedance.\*\*

**Output Energy Sequence:** Multiple levels, configurable from 150 joules to 360 joules (200 joules min for Japan). Factory default settings of 200J, 300J, 360J.

**Output Energy Accuracy:** ±10% into 50 ohms, ±15% into 25 to 100 ohms.

**Shock Advisory System:** An ECG analysis system that advises whether a shock is appropriate; meets rhythm recognition criteria specified in DF39.

The device charges for shock only when the Shock Advisory System advises defibrillation.

#### Device Capacity:

**Typical:** Thirty (30) full discharges or 210 minutes of "on time" with a fully charged device.

**Minimum:** Twenty (20) full discharges or 140 minutes of "on time" with a fully charged device.


**Shock Charge Time:** Charge times with a fully charged device: 200 joules in less than 9 seconds, 360 joules in less than 15 seconds.

**System Recharge Times:** Recharge times with a fully discharged device: Able to deliver 6 shocks or provide 42 minutes of operating time after 24 hours of recharge time and 20 shocks or 140 minutes of operating time after 72 hours of recharge time with a new CHARGE-PAK at temperatures above 15° C (59° F).

#### Controls:

Lid Release/ON-OFF—Controls device power.

SHOCK button (semi-automatic version)—delivers defibrillation energy. After electrodes are attached to a patient, the fully automatic version of the device delivers a shock, if appropriate, not requiring operator intervention.

**Electrical Protection:** Input protected against high voltage defibrillator pulses per  IEC60601-1/EN60601-1.

**Safety Classification:** Internally powered equipment. IEC60601-1/EN60601-1.

### PHYSICAL CHARACTERISTICS

**Height:** 10.7 cm (4.2 in).

**Width:** 20.3 cm (8.0 in).

**Depth:** 24.1 cm (9.5 in), excluding handle.

**Weight:** 2.0 kg (4.5 lb) with CHARGE-PAK and electrodes

### USER INTERFACE

**User Interface:** The user interface includes voice prompts, audible tones and graphic prompts.

**Readiness Display:** The readiness display shows the device status.

**OK Indicator:** Shows "OK" when the last self-test was completed successfully. When the "OK" indicator is visible, all other indicators are not visible. The "OK" indicator is not displayed during device operation.

**CHARGE-PAK Indicator:** When displayed, replace the CHARGE-PAK™ battery charger.

**Attention Indicator:** When first displayed, at least six (6) discharges or 42 minutes of operating time remain.

**Service Indicator:** Service required when displayed.

### ENVIRONMENTAL

**Note:** All performance specifications defined assume the unit has been stored (two hours minimum) at operating temperature prior to operation.

**Operating Temperature:** 0° to +50° C (+32° to +122° F).

**Storage Temperature:** -40° to +70° C (-40° to +158° F) with CHARGE-PAK and electrodes, maximum exposure time limited to one week.

**Atmospheric Pressure:** 760 mmHg to 429 mmHg, 0 to 15,000 feet above sea level.

**Relative Humidity:** 5 to 95% (non-condensing).

**Water Resistance:** IEC60529/EN60529 IPX4 "Splash proof" with electrodes connected, CHARGE-PAK installed.

**Shock:** MIL-STD-810E, Method 516.4, Procedure 1, (40g, 6-9 ms pulse, ½ sine each axis).

**Vibration:** MIL-STD-810E, Method 514.4, Helicopter - category 6 (3.75 Grms) and Ground Mobile - category 8 (2.85 Grms).

### DEFAULT SETTINGS

**Energy Sequence:** Energy sequence is set to 200J, 300J, 360J.

**Motion Detection:** The motion detection system is set to on during analysis.

**Energy Protocol:** The energy protocol is set to increase energy only after a lower energy shock was unsuccessful.

**Stack Shocks:** Stack shocks option is set to off.

**Turn-On Prompt:** The turn-on prompt is set to provide voice prompts upon power on.

**CPR Time:** The CPR Time is set to 120 seconds.

**Voice Prompt Volume:** The voice prompt volume is set to high.

### ACCESSORIES

#### CHARGE-PAK Battery Charger

**Type:** Li/SO<sub>2</sub>Cl<sub>2</sub> Lithium Sulfuryl Chloride, 11.7V, 1.4 amp-hours.

**Replacement:** Replace the CHARGE-PAK battery charger and QUIK-PAK™ electrodes packet after using the defibrillator, if the CHARGE-PAK symbol appears in the readiness display or when the Use By date is reached (typically 2 years).

**Weight:** 80.5 grams (0.18 lb).

#### QUIK-PAK Electrode Pads

**Pads:** ECG is received from disposable defibrillation electrodes, standard placement (anterior-lateral).

**Pads Packaging:** User intuitive, rapid release QUIK-PAK electrodes allow the electrode pads to be preconnected to the device and protected under a top cover.

**Pads Replacement:** Replace every two (2) years.

#### Infant/Child Reduced Energy Defibrillation

**Electrodes:** For use on infants and children less than 8 years of age or less than 55 lbs (25kg).

### DATA STORAGE

**Memory Type:** Internal digital memory.

**ECG Storage:** Dual patient data storage. Minimum 20 minutes of ECG stored for the current patient, summarized data stored for the previous patient.

#### Report Types:

- Continuous ECG – A continuous patient ECG report.
- Continuous Summary report – A summary of critical resuscitation events and ECG waveform segments associated with these events.
- Event Log report – A report of time stamped markers, which reflect operator and device activity.
- Test Log report – A device self-test activity report.

**Capacity:** Minimum 200 time-stamped event log markers.

**Communications:** Wireless transfer to a personal computer.

**Data Review:** Physio-Control provides an array of tools to meet customer needs for data viewing and analysis.

\*\* The specifications apply from 25 to 200 ohms. Voltage compensation is limited to the voltage that would result in delivery of 360 joules into 50 ohms.

\* Stiell IG, et al. *Circulation* 2007;115:1511-1517. All claims valid as of October 2012.

All specifications are at 20° C unless otherwise stated.



**Physio-Control Headquarters**  
Redmond, WA 98052  
www.physio-control.com

**Customer Support**  
P. O. Box 97006  
Redmond, WA 98073  
Toll free 800 442 1142  
Fax 800 426 8049

**Physio-Control Canada**  
Physio-Control Canada Sales, Ltd.  
7111 Syntex Drive, 3rd Floor  
Mississauga, ON  
L5N 8C3  
Canada  
Toll free 800 895 5896  
Fax 866 430 6115



Physio-Control, Inc., 11811 Willows Road NE, Redmond, WA 98052 USA